

Zertifikat

Prüfungsbestätigung für Nachweis nach NEV/VEMV und EU/EWR

CERTIFEL ZA1

Zertifikat Ref.Nr.:

EL-1727

Seite 1 von 1

Produkt	Medizinkanal
Auftraggeber	EHB Brüstungstechnik AG Alte Strasse 11, CH-4665 Oftringen
Hersteller	EHB Brüstungstechnik AG Alte Strasse 11, CH-4665 Oftringen
Fertigungsstätte	EHB Brüstungstechnik AG Alte Strasse 11, CH-4665 Oftringen
Handelsmarke	EHB
Typ/Modell	SPI 02
Nenndaten	230V~, 50Hz, 13A Class I, IP40
Normen Sicherheitstechnik	EN 60601-1:90 + A1:93 + A2:95 EN ISO 11197:09 IEC 60601-1(ed.2);am1;am2
Normen EMV	EN 60601-1-2:07 EN ISO 11197:09 IEC 60601-1-2(ed.3)
Zusätzliche Angaben	---



Das Produkt wurde nach den oben aufgeführten normativen Dokumenten geprüft. Es kann angenommen werden, dass damit der Nachweis erbracht wurde um den Anforderungen der Schweizer Behörden (NEV/VEMV) und den Europäischen Direktiven (EU/EWR) zu genügen.

Die Prüfergebnisse sind in folgenden Bericht-Nummern festgehalten: 11-EL-0054.01 + .02 + .10(EMV)



Martin Plüss
Zertifizierung Produkte



Fehraltorf, 2011-12-07

Luppenstrasse 1
CH-8320 Fehraltorf

Tel. +41 44 956 11 11
Fax +41 44 956 11 22
info@electrosuisse.ch
www.electrosuisse.ch

Test Report issued under the responsibility of:



**EXPERT REPORT
EN ISO 11197 (2004)
Medical supply units**

Report reference No......: 11-EL-0098.01
Date of issue: 2011-10-10
Total number of pages: 17

CB Testing Laboratory.....: Electrosuisse
Address: Luppmenstrasse 1, CH-8320 Fehraltorf
Switzerland



Applicant's name.....: EHB Brüstungstechnik
Address: Alte Strasse 11, 4665 Oftringen

Test specification:

Standard.....: EN ISO 11197: (2004)
Test procedure: Testing of the following sub clauses :
5, 6,21 to 28,54,55,59
Non-standard test method.....: N/A

Test Report Form No......: TRF_EN ISO11197-1A
Test Report Form(s) Originator: Electrosuisse
Master TRF.....: Dated: 11-07

Copyright © Electrosuisse, Luppmenstrasse 1, 8320 Fehraltorf, Switzerland. All rights reserved.

This publication may be reproduced in whole or in part for non-commercial purposes as long as the Electrosuisse is acknowledged as copyright owner and source of the material. Electrosuisse takes no responsibility for and will not assume liability for damages resulting from the reader's interpretation of the reproduced material due to its placement and context.

This report is not valid as a Test Report unless signed by an approved Testing Laboratory and appended to a Test Certificate issued by an NCB.

Test item description: Medical supply unit
Trade Mark: EHB
Manufacturer: EHB Brüstungstechnik, Alte Strasse 11, 4665 Oftringen
Model/Type reference: SPI01 GU / Typ-Nr. 8077.2.013 and
SPI01 GO/ Typ-Nr. 8077.2.014
Ratings: -

EMC Test Report

Number, Revision: PB 11-EL-0054

Date: November 29, 2011

Client: EHB Brüstungstechnik AG
Alte Strasse 11
4665 Oftringen

Equipment under Test: Medical supply unit

Magnitude of Test: EMC-Tests according to harmonized standards with respect to Directive 93/42/EEC concerning medical devices:



- EN 60601-1-2:2007
- EN ISO 11197 (relevant parts for EMC)

Result of Test: The equipment under test (EUT) conforms to the requirements mentioned above.

Author: Rudolf Hasler

Telephone: +41(0)58 252 4282

Distribution List: Client (Original), PST2, Archive

Function	Department	Name	Signature	Date
Test engineer	EAL	Rudolf Hasler		29.11.11
Head of Albislab	EAL	Urs von Känel		30.11.2011

The client named in this report has the right to reproduce and quote this report as necessary, in connection with the consent of the equipment under test.

The original test report is authorized with a signature on the front page by the responsible persons of the test laboratory and must be archived for the time stated by law by the client named in this report. Only the original test report is valid for legal purposes.

This test report shall not be reproduced except in full without the expressed written consent of Electrosuisse